T NO 124-2 PREBID QUERIES

Firm Name	Firm Request	Tender Document	User department Remarks
Fresenius Kabi India Pvt. Ltd.	Leuko Deplition Filter Sets : Original Specification 1. Filter should be able to Leuko-deplete red cell from leukocyte contamination separately for one-unit red cell Amendment Specification 1. Filter should be bedside able to Leuko- deplete red cell from leukocyte contamination separately for one-unit red cell	Specification for Leuko Deplition Filter Sets Eilter should be able to Leuko- deplete red cell from leukocyte contamination separately for one- unit red cell .	PTHHP Remarks: The amendment request of the Leuko Deplition Filter Sets specification is accepted and for the patient convenience bedside Leuko Deplition Filter Sets are needed. Hence Item Read as Leuko Deplition Filter Bed side.
	SUPPLY CONDITIONS Original:- Supplies against a purchase order shall be completed within 60 days otherwise liquidated damages are levied by APMSIDC as mentioned in clause 18.1. Amendment:- Supplies against a purchase order shall be completed within 90days otherwise liquidated damages are levied by APMSIDC as mentioned in clause 18.1.	Supplies against a purchase order shall be completed within 60 days otherwise liquidated damages are levied by APMSIDC as mentioned in clause 18.1.	Not Considered.